

IP4B

Capacitive Coupling Clamp for EFT / Burst according to IEC/EN 61000-4-4 Ed. 3

OVERVIEW

The IP4B capacitive coupling clamp is used to transfer an interference signal, using capacitive coupling onto control or data lines. The IP4B can be used with the HAEFELY HIPOTRONICS EFT/Burst Immunity Test System AXOS 5, AXOS 8 or PEFT 8010.

The IP4B capacitive coupling clamp is built according to the standard IEC / EN 61000-4-4 Edition 3.

Please also refer to this standard also for exact test setup.



FEATURES & BENEFITS

- ☑ Designed as per IEC 61000-4-4 Ed.3
- ☑ EFT/Burst testing of signal and data lines
- ☑ Active coupling length 1 m

APPLICATION

The IP4B capacitive coupling clamp is primarily used to inject fast transient and burst interference pulses into signal and data cables, i.e. into any type of connection to peripheral equipment. The IEC standard 61000-4-4 Ed. 3 also allows the capacitive coupling method to be used for pulse injection into ac and dc power lines when no appropriate decoupling network is available The coupling capacitance (typically 100pF) between the coupling clamp and the cable inserted depends on the cable type, the diameter and other factors like screening, etc.

Europe

Haefely Test AG Birsstrasse 300 4052 Basel Switzerland

★ 41 61 373 4111
 ★ 41 61 373 4912
 ★ sales@haefely.com

2 + 86 10 8578 8099
 3 + 86 10 8578 9908
 3 sales@haefely.com.cn

China

Haefely Test AG Representative Beijing Office

8-1-602, Fortune Street
No. 67, Chaoyang Road, Chaoyang District

Beijing, China 100025

CALIBRATION OF IP4B

The transducer plate shall be placed in the capacitive coupling clamp such that the end with the connection is aligned with the end of the lower coupling plate. The connecting adapter shall support a low impedance connection to ground reference plane for grounding of the 50Ω coaxial measurement terminator/attenuator. The distance between the transducer plate and the $50~\Omega$ measurement terminator/attenuator shall not exceed 0,1m.

OPTIONS AND ACCESSORIES

EFT Verification Set (includes 50 Ω and 1 k Ω attenuator) Article no. 2499951

PAT 50A attenuator 54 dB, 50 Ω Article no. 2495491

PAT 1000 attenuator 60 dB, 1 k Ω Article no. 2495441

Transducer Plate for IP4B *Article no. 0781820*

Accredited Calibration according to ISO 17025 Article no. 2490955

North America

Hipotronics, Inc. 1650 Route 22 N Brewster, NY 10509 United States

2 + 1 845 279 3644
 3 + 1 845 279 2467
 3 sales@hipotronics.com

HAEFELY HIPOTRONICS has a policy of continuous product improvement. Therefore we reserve the right to change design and specification without notice.







OFFICES:

Europe

Haefely Test AG Birsstrasse 300 4052 Basel Switzerland

***** + 41 61 373 4111

+ 41 61 373 4912

sales@haefely.com

China

Haefely Test AG Representative Beijing Office 8-1-602, Fortune Street No. 67, Chaoyang Road, Chaoyang District Beijing, China 100025

> + 86 10 8578 8099 + 86 10 8578 9908

sales@haefely.com.cn

North America

Hipotronics, Inc. 1650 Route 22 N Brewster, NY 10509 United States

+ 1 845 279 3644 + 1 845 279 2467 sales@hipotronics.com